



Health  
Canada

Santé  
Canada

## Notice

November 3, 2015

Our reference number: 15-111988-348

### **Re: Health Canada Requirements for Manufacturers Relating to Manufacturing Change to Tyvek® 1073B and 1059B**

Health Canada is aware that DuPont has altered their manufacturing process for two grades of Tyvek (1073B and 1059B), which they are referring to as “Transition Tyvek”. These materials are extensively used in sterile medical device packaging and this change will impact a large number of licensed medical devices.

As indicated in the May 7, 2014 notice (see Appendix) relating to the regulatory approach being taken by Health Canada in this situation, Health Canada is requiring attestations from current Class III and IV Medical Device Licence Holders, regardless of whether or not Tyvek is present in the product. Attestations are not required for Class II licences at this time.

To assist with this attestation, packages for all manufacturers of Class III and IV devices were e-mailed to their authorized regulatory contact on September 30, 2015. Although the packages have been e-mailed to the regulatory contact, the responsibility for signing the attestation and verifying the information in the “September 30, 2015 Licence Notification” will be that of the medical device licence holders (manufacturers).

A complete “September 30, 2015 Licence Notification” form and signed “Attestation” will need to be submitted by December 31, 2015. Instructions on how to complete the form have been provided in the message to the regulatory contacts. Manufacturers are encouraged to contact their regulatory contacts to ensure that this attestation is completed and submitted to Health Canada by the due date.

Please direct any questions or comments regarding the content of this Notice to the following:

Device Licensing Services Division  
Medical Devices Bureau  
Therapeutic Products Directorate  
Health Products and Foods Branch  
Health Canada  
2934 Baseline Road, Tower B  
Address Locator: 3403A  
Ottawa, Ontario  
K1A 0K9

Email: [device\\_licensing@hc-sc.gc.ca](mailto:device_licensing@hc-sc.gc.ca)

Telephone: 613-957-7285

Fax Number: 613-957-6345

**Canada** 

## Appendix

### Notice to Sterile Medical Device Manufacturers

#### Health Canada Requirements for Manufacturers Relating to Manufacturing Change to Tyvek® 1073B and 1059B

May 7, 2014

##### **Purpose**

Health Canada has been notified that DuPont is altering their manufacturing process for two grades of Tyvek® (1073B and 1059B). These materials are extensively used in sterile medical device packaging and this change will impact a large number of licenced medical devices.

This notice is intended to communicate Health Canada's approach to managing the change in Tyvek®'s manufacturing process and Health Canada's expectations from manufacturers, including requirements for a modified process for filing a significant change amendment for Class III and IV licences.

##### **Scope**

This notice is relevant to any manufacturers of sterile medical devices who hold active Medical Device Licences that use or will use DuPont's new Tyvek® material in the packaging of their currently licenced devices.

Although significant change amendments are only required for Class III and IV medical device licences, the risk assessment and risk management principles described in this notice should be applied to Class II sterile devices to ensure they continue to meet the safety and effectiveness requirements of the *Medical Devices Regulations*.

##### **Health Canada's Approach to Managing the Change in Tyvek®**

DuPont is currently undertaking tests to evaluate characteristics of the new Tyvek® material in a variety of packaging configurations and sterilization processes.

Health Canada will be accepting DuPont's test results as a pilot Medical Device Master File and, in cooperation with its international counterparts, will be reviewing the data.

DuPont will be making relevant testing summaries and/or data available to medical device manufacturers.

**IMPORTANT NOTES:**

- In a number of locations in this notice, reference is made to notification and attestations to be provided to Health Canada. Health Canada will be sending specific communications to each active licence holder. **No notification is required prior to receipt of this communication.**
- No notification is required for Class II licences.
- The approach described here assumes positive outcomes following the review of DuPont's Tyvek® Medical Device Master File. Should issues be identified through the review process, changes to the approach may occur.

For Class III and IV licences, the regulatory requirement for an amendment will be satisfied through a modified process. Specifically, following the review of DuPont's Tyvek® Medical Device Master File, Health Canada will require notification from manufacturers identifying the licences impacted by the change and attesting the configurations and sterilization methods are represented by the testing conducted by DuPont. Standard licence amendment applications may be required in cases where DuPont's testing is not adequately reflective of the package configuration - further details are provided below.

This approach is being taken due to the unique nature and breadth of the use of Tyvek® in medical device packaging and should be considered unique to this case.

***Manufacturer's Obligations - Class III and IV Licence Holders***

Manufacturers are responsible for the safety of the devices they sell and any changes made to them.

Manufacturers are expected to review the information provided to them by DuPont and assess if it is applicable to the packaging configuration and sterilization process of their device.

- If the DuPont evidence is clearly applicable to their devices, the manufacturer should notify Health Canada, identifying the licences impacted by the change and attesting the configurations and sterilization methods are represented by the testing conducted by DuPont.
- If the packaging configuration and/or sterilization process is not covered by the DuPont testing, the manufacturer is to assess the differences and determine if material or process changes may be reasonably expected to negatively impact device sterility over the stated shelf-life.
  - If the assessment concludes that introducing the new sterile barrier will not significantly impact the packaging's performance, the manufacturer should notify Health Canada, identifying the licences impacted by the change and attesting that the manufacturer's configurations and sterilization methods are sufficiently represented by the testing conducted by DuPont.
  - If there is a reasonable expectation that differences may negatively impact the device sterility over the stated shelf-life, the manufacturer is to submit a Medical Device Licence Amendment Application for a manufacturing change and provide

supporting validation using their packaging configuration and sterilization process. This may include sterilization validation, shelf-life validation and package integrity assessments.

- All documentation related to the assessments described above should be available upon request.

### ***Manufacturer's Obligations - Class II Licence Holders***

Manufacturers are responsible for the safety of the devices they sell and any changes made to them.

Although notification or licence amendments are only required for Class III and IV medical device licences, the risk assessment and risk management principles described in this notice should be applied for Class II sterile devices to ensure the Class II devices continue to meet the safety and effectiveness requirements of the *Medical Devices Regulations*.

### ***New Licence Applications for Devices using the new Tyvek®***

Requirements for New Medical Device Licence Applications will remain unchanged. Support for the packaging will be required, the form of such support will vary based on the specific circumstances, but may involve reference to DuPont's Tyvek® Medical Device Master File, existing licences, supportive testing, or a detailed scientific justification.

### **Contact**

If you have any questions relating to this notice, please contact the Medical Devices Bureau at [mdb\\_enquiries@hc-sc.gc.ca](mailto:mdb_enquiries@hc-sc.gc.ca) or 613-957-7285.